



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM ADAMTS-13 INH

REF 5450451
LOT RP39B0S.01
 30 - APR - 2015

| Inhalt Content | Reagenz Reagent | LOT | |
|-------------------|---------------------------|---|-------------|
| 1 x 6 | Adamts13 INH Plate | RPS37BS | 31-AUG-2015 |
| 1 x 0.3mL | Adamts13 INH Conjugate | RPV31B0 | 30-APR-2015 |
| 1x 80mL | Wash buffer Concentrate | RNA34B0 | 30-SEP-2015 |
| 1 x 90mL | Incubation buffer | RNB33B0 | 30-SEP-2015 |
| 1 x 12mL | TMB Substrate | RKN33B0 | 30-SEP-2015 |
| 1 x 12mL | Stop solution | RKK33B0 | 31-OCT-2015 |
| 1 x 0,5mL | Adamts13 INH Calibrator 1 | 1PT31B0 | 30-JUN-2015 |
| 1 x 0.5mL | Adamts13 INH Calibrator 2 | 2PT31B0 | 30-JUN-2015 |
| 1 x 0.5mL | Adamts13 INH Calibrator 3 | 3PT32B0 | 31-JUL-2015 |
| 1 x 0.5mL | Adamts13 INH Calibrator 4 | 4PT31B0 | 31-JUL-2015 |
| 1 x 0.5mL | Adamts13 INH Calibrator 5 | 5PT31B0 | 31-JUL-2015 |
| 1 x 0.5mL | Adamts13 INH High Control | RPY31B0 | 31-JUL-2015 |
| 1 x 0.5mL | Adamts13 INH Low Control | RPZ31B0 | 31-JUL-2015 |

| Analytische Prozedur Analytical Procedure | Anforderung Requirement | Ergebnis Result | |
|--|----------------------------|--------------------|--------------|
| OD | Calibrator 1 | > 1.0 | 1.364 OK |
| | Calibrator 5 | < 0.1 | 0.003 OK |
| Verwendbarkeit Usability | High Control (RPY31B0) | 53.1 – 79.6 | 54.486 OK |
| | Low Control (RPZ31B0) | 4.21 – 9.83 | 4.220 OK |

Durchführung von
Prepared by

Michael RIHA

20-Nov-2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER

20-Nov-2013
Datum / Date: