



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM ADAMTS-13 INH

REF 5450451

LOT RP43B0S.02

⌚ 28 - FEB - 2015

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 6	Adamts13 INH Plate	RPS36BS	31-JUL-2015
1 x 0.3mL	Adamts13 INH Conjugate	RPV31B0	30-APR-2015
1x 80mL	Wash buffer Concentrate	RNA32B0	28-FEB-2015
1 x 90mL	Incubation buffer	RNB32B0	30-APR-2015
1 x 12mL	TMB Substrate	RKN32B0	31-MAY-2015
1 x 12mL	Stop solution	RKK32B0	30-JUN-2015
1 x 0,5mL	Adamts13 INH Calibrator 1	1PT31B0	30-JUN-2015
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B0	30-JUN-2015
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT31B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH High Control	RPY31B0	31-JUL-2015
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B0	31-JUL-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.074	OK
	Calibrator 5	< 0.1	0.006	OK
Verwendbarkeit Usability	High Control (RPY31B0)	53.1 – 79.6	58.763	OK
	Low Control (RPZ31B0)	4.21 – 9.83	9.395	OK

Durchführung von
Prepared by

Michael RIHA

3-Mär-2014

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

3-Mär-2014

Datum / Date: