

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRP00AZ
---	---	----------

### TECHNOZYM ADAMTS-13 INH

**REF**      5450451  
**LOT**      RP4AB0S.01  
      30 - Jun - 2015

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 6	Adamts13 INH Plate	RPS47BS	30-Apr-2016
1 x 0.3mL	Adamts13 INH Conjugate	RPV42B0	30-Apr-2016
1x 80mL	Wash buffer Concentrate	RNA41B0	29-Feb-2016
1 x 90mL	Incubation buffer	RNB42B0	30-Apr-2016
1 x 12mL	TMB Substrate	RKK41B0	31-Mar-2016
1 x 12mL	Stop solution	RKN34B0	30-Nov-2015
1 x 0,5mL	Adamts13 INH Calibrator 1	1PT31B0	30-Jun-2015
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B0	30-Jun-2015
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B0	31-Jul-2015
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B0	31-Jul-2015
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT31B0	31-Jul-2015
1 x 0.5mL	Adamts13 INH High Control	RPY31B0	31-Jul-2015
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.315	OK
	Calibrator 5	< 0.1	0.004	OK
Verwendbarkeit Usability	High Control (RPY31B0)	53.1 – 79.6	68.11	OK
	Low Control (RPZ31B0)	4.21 – 9.83	8.60	OK

**Durchführung von  
Prepared by**

Michael RIHA 

28-Jul-2014

Datum / Date:

**Freigabe von  
Approved by**

Fikret NASUFI 

28-Jul-2014

Datum / Date: