



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRP0SAZ

TECHNOZYM ADAMTS-13 INH ELISA Kit 48T

REF 5450451

LOT RP52B0S.01

⌚ 31-Aug-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 6	Adamts13 INH Plate	RPS4ABS	31-Aug-2016
1 x 0.3mL	Adamts13 INH Conjugate	RPV43B0	31-Oct-2016
1x 80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1 x 90mL	Incubation buffer	RNB51B0	29-Feb-2017
1 x 12mL	TMB Substrate	RKN41B0	31-Aug-2016
1 x 12mL	Stop solution	RKK43B0	31-Dec-2016
1 x 0,5mL	Adamts13 INH Calibrator 1	1PT31B1	31-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B1	31-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B1	31-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B1	31-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT31B1	31-Nov-2016
1 x 0.5mL	Adamts13 INH High Control	RPY31B1	31-Nov-2016
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B1	31-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.47	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY31B1)	53.1 – 79.64 U/mL	66.05	OK
	Low Control (RPZ31B1)	4.21 – 9.83 U/mL	8.28	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

14-Apr-2015
Datum / Date:

Freigabe von
Approved by

Lavdim NASUFI *Lavdim Nasufi*

14-Apr-2015
Datum / Date: