
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP0SAZ
---	---	----------

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451

LOT RP57B0S.01

 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 6	Adamts13 INH Plate	RPS54BS	31-Dec-2017
1 x 0.3mL	Adamts13 INH Conjugate	RPV51B0	31-May-2017
1x 80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1 x 90mL	Incubation buffer	RNB52B0	31-Jul-2017
1 x 12mL	TMB Substrate	RKN52B0	31-Dec-2017
1 x 12mL	Stop solution	RKK52B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT52B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH High Control	RPY31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.50	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RPY31B1)	53.10 – 79.64 U/mL	75.26	OK
	Low Control (RPZ31B1)	4.21 – 9.83 U/mL	8.56	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasefi*

23-Dec-2015
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

23-Dec-2015
Datum / Date: