




**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**


DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451

LOT RP58B0S.01

 31-May-2017

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 INH Plate	RPS54BS	31-Dec-2017
1 x 0.3mL	Adamts13 INH Conjugate	RPV51B0	31-May-2017
1x 80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1 x 90mL	Incubation buffer	RNB52B0	31-Jul-2017
1 x 12mL	TMB Substrate	RKN52B0	31-Dec-2017
1 x 12mL	Stop solution	RKK52B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH High Control	RPY51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B0	31-Dec-2017


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.44	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	70.73	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	11.02	OK

**Durchführung von
Prepared by**

Michael KAFKA 

21-JAN-2016
Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

21-JAN-2016
Datum / Date: