

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6PT0AZ.02
---	---	-------------

TECHNOZYM ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461
LOT 6PT31B0.01
 30 - Jun - 2015

Inhalt Content	Reagenz Reagent	LOT	
1X0.5mL	Adamts-13 INH Calibrator 1	1PT31B0	30-Jun-2015
1X0.5mL	Adamts-13 INH Calibrator 2	2PT31B0	30-Jun-2015
1X0.5mL	Adamts-13 INH Calibrator 3	3PT31B0	31-Jul-2015
1X0.5mL	Adamts-13 INH Calibrator 4	4PT11B0	31-Jul-2015
1X0.5mL	Adamts-13 INH Calibrator 5	5PT31B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.27	OK
	Calibrator 5	< 0.1	0.012	OK
Verwendbarkeit Usability (tested with Adamts-13 INH Plate RPS36B0)	High Control (RPY31B0)	53.1 – 79.64 %	61.52	OK
	Low Control (RPZ31B0)	5.62 – 8.42 %	8.4	OK

Durchführung von **Fikret NASUFI**
Prepared by

F. Nasufi

16-Sep-2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER
[Signature]

16. SEP. 2013
Datum / Date: