



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQ6PT0AZ.02

**TECHNOZYM ADAMTS-13 INH Calibrator Set 5x0.5mL**

**REF** 5450461

**LOT** 6PT31B0.02

30 - Jun - 2015

Inhalt Content	Reagenz Reagent	LOT	
1X0.5mL	Adamts-13 INH Calibrator 1	1PT31B0	30-Jun-2015
1X0.5mL	Adamts-13 INH Calibrator 2	2PT31B0	30-Jun-2015
1X0.5mL	Adamts-13 INH Calibrator 3	3PT31B0	31-Jul-2015
1X0.5mL	Adamts-13 INH Calibrator 4	4PT11B0	31-Jul-2015
1X0.5mL	Adamts-13 INH Calibrator 5	5PT31B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.220	OK
	Calibrator 5	< 0.1	0.019	OK
Verwendbarkeit Usability (tested with Adamts-13 INH Plate RPS36B0)	High Control (RPY31B0)	53.1 – 79.64 %	67.83	OK
	Low Control (RPZ31B0)	4.21 – 9.83 %	9.22	OK

**Durchführung von  
Prepared by**

Michael RIHA

31-MAR-2014  
Datum / Date:

**Freigabe von  
Approved by**

FF"

31-MAR-2014  
Datum / Date: