



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ6PT0AZ.02

TECHNOZYM ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461

LOT 6PT31B0.03

🕒 30 - Jun - 2015

Inhalt Content	Reagenz Reagent	LOT	🕒
1X0.5mL	Adamts-13 INH Calibrator 1	1PT31B0	30-Jun-2015
1X0.5mL	Adamts-13 INH Calibrator 2	2PT31B0	30-Jun-2015
1X0.5mL	Adamts-13 INH Calibrator 3	3PT32B0	31-Jul-2015
1X0.5mL	Adamts-13 INH Calibrator 4	4PT31B0	31-Jul-2015
1X0.5mL	Adamts-13 INH Calibrator 5	5PT31B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.33	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability (tested with Adamts-13 INH Plate RPS49B0)	High Control (RPY31B0)	53.1 – 79.64 U/mL	66.47	OK
	Low Control (RPZ31B0)	4.2 – 9.83 U/mL	8.57	OK

Durchführung von
Prepared by

Fikret NASUFI

09-Sep-2014
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

09-Sep-2014
Datum / Date: