



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6PT00AZ

TECHNOZYM ADAMTS-13 INH Calibrator Set

REF 5450461

LOT 6PT41B0.01

⌚ 30-Nov-2016

| Inhalt Content | Reagenz Reagent | LOT | ⌚ |
|-------------------|---------------------------|---------|-------------|
| 1 x 0,5mL | Adamts13 INH Calibrator 1 | 1PT31B1 | 30-Nov-2016 |
| 1 x 0.5mL | Adamts13 INH Calibrator 2 | 2PT31B1 | 30-Nov-2016 |
| 1 x 0.5mL | Adamts13 INH Calibrator 3 | 3PT32B1 | 30-Nov-2016 |
| 1 x 0.5mL | Adamts13 INH Calibrator 4 | 4PT31B1 | 30-Nov-2016 |
| 1 x 0.5mL | Adamts13 INH Calibrator 5 | 5PT31B1 | 30-Nov-2016 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| OD | Calibrator 1 | > 1.0 | 1.20 | OK |
| | Calibrator 5 | < 0.1 | 0.01 | OK |
| Verwendbarkeit Usability | High Control (RPY31B1) | 53.10 – 79.64 U/mL | 68.50 | OK |
| | Low Control (RPZ31B1) | 4.21 – 9.83 U/mL | 6.62 | OK |

Durchführung von
Prepared by

Michael RIHA 

02-FEB-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

02-Feb-2015
Datum / Date: