

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6PT00AZ
---	---	-----------

### TECHNOZYM ADAMTS-13 INH Calibrator Set 5x0.5mL

**REF**      5450461  
**LOT**      6PT41B0.02  
      30-Nov-2016

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT32B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.22	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RPY31B1)	53.10 – 79.64 U/mL	65.33	OK
	Low Control (RPZ31B1)	4.21 – 9.83 U/mL	8.91	OK

**Durchführung von  
Prepared by**

*Fikret NASUEI*

28-Sep-2015  
Datum / Date:

**Freigabe von  
Approved by**

**Michael KAFKA** *[Signature]*

28-Sep-2015  
Datum / Date: