

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6PT00AZ
-----------------------------------------------------------------------------------	-------------------------------------------------------	-----------

## TECHNOZYM® ADAMTS-13 INH Calibrator Set

REF      5450461

LOT      6PT51B0.01

      31-Dec-2017

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.46	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	70.76	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	10.68	OK

**Durchführung von**  
**Prepared by**

Michael KAFKA 

21-JAN-2016  
Datum / Date:

**Freigabe von**  
**Approved by**

Fikret NASUFI 

21-JAN-2016  
Datum / Date: