

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT0AZ.02
---	---	-------------

TECHNOZYM ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463
LOT 7PT31B0.01
 31 - Jul - 2015

Inhalt Content	Reagenz Reagent	LOT	
1X0.5mL	Adamts-13 INH High Control	RPY31B0	31-Jul-2015
1X0.5mL	Adamts-13 INH Low Control	RPZ31B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability (tested with Adamts-13 INH Plate RPS36BS)	High Control (RPY31B0)	53.1 – 79.64 %	67.50	OK
	Low Control (RPZ31B0)	5.62 – 8.42 %	8.23	OK

Durchführung von **Fikret NASUFI**
Prepared by

F. Nasufi

16-Sep-2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER
Nikolaus BINDER

18. SEP. 2013
Datum / Date: