



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7PT00AZ

TECHNOZYM ADAMTS-13 INH Control Set

REF 5450463
LOT 7PT51B0.01
⌚ 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 0.5mL	Adamts13 INH High Control	RPY31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.20	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY31B1)	53.10 – 79.64 U/mL	68.50	OK
	Low Control (RPZ31B1)	4.21 – 9.83 U/mL	6.62	OK

Durchführung von
Prepared by

Michael RIHA *Michael Riha*

02-FEB-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI *Fikret Nasufi*

02-Feb-2015
Datum / Date: