

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT00AZ
---	---	-----------

TECHNOZYM ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463
LOT 7PT51B0.02
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY31B1	30-Nov-2016
1 x 0.5mL	Adamts13 INH Low Control	RPZ31B1	30-Nov-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.23	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY31B1)	53.10 – 79.64 U/mL	67.25	OK
	Low Control (RPZ31B1)	4.21 – 9.83 U/mL	8.36	OK


**Durchführung von
Prepared by**



Fikret NASUE

14-Sep-2015
Datum / Date:

**Freigabe von
Approved by**



14-Sep-2015
Datum / Date: