

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT00AZ
---	---	-----------

TECHNOZYM® ADAMTS-13 INH Control Set

REF 5450463
LOT 7PT52B0.01
 31-Dec-2017

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.46	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	70.76	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	10.68	OK

**Durchführung von
Prepared by**

Michael KAFKA 

21-JAN-2016
Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

21-JAN-2016
Datum / Date: