



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRQ00AZ

## TECHNOZYM ADAMTS-13 ELISA Kit 2x48T.

REF            5450501

LOT            RQ33B00.02


🕒            31-Mar-2015

Inhalt Content	Reagenz Reagent	LOT	🕒
1x48T	Adamts13 Plate	RQS33B0	30-Jun-2015
1x0.3mL	Adamts13 Conjugate	RQV32B0	31-Aug-2015
1x3mL	AdamTS13 Activity Substrate	RQW31B0	31-Mar-2015
1x6mL	AdamTS13 Antigen Substrate	RQX32B0	31-Mar-2015
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL32B0	31-Mar-2015
1x6mL	AdamTS13 Antigen Stop Solution	RQK32B0	31-Mar-2015
1x80mL	Wash buffer Concentrate	RNA33B0	31-May-2015
1x90mL	Incubations buffer	RNB32B0	30-Apr-2015
1x0,5ml	Adamts 13 Calibrator 1	1UT32B0	30-Jun-2015
1x0,5ml	Adamts 13 Calibrator 2	2UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 3	3UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 4	4UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 5	6UT32B0	30-Jun-2015
1x0,5ml	Adamts 13 High Control	RUY32B0	31-Jul-2015
1x0,5ml	Adamts 13 Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RUY32B0 activity	46.3– 77.2 %	66.7    OK
	High control RUY32B0 antigen	0.63 – 1.05 µg/mL	0.83    OK
	Low control RUZ32B0 activity	14.0 – 23.4 %	20.4    OK
	Low control RUZ32B0 antigen	0.19 – 0.32 µg/mL	0.26    OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	>250	500.4    OK
	Gain: 60 Antigen ΔRFU Cal1-Cal5	>250	413    OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included	Ja / Yes	Ja / Yes	OK

Durchführung von  
Prepared by

Fikret NASUFI




---

27-Sep-2013  
Datum / Date:

Freigabe von  
Approved by

Nikolaus BINDER



---

30. SEP. 2013  
Datum / Date: