



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRQ00AZ

TECHNOZYM ADAMTS-13 ELISA Kit 2x48T.

REF 5450501

LOT RQ34B00.01

⌚ 28 - Feb - 2015

Inhalt Content	Reagenz Reagent	LOT	⌚
1x48T	Adamts13 Plate	RQS34B0	30-Sep-2015
1x0.3mL	Adamts13 Conjugate	RQV31B0	28-Feb-2015
1x3mL	AdamTS13 Activity Substrate	RQW31B0	31-Mar-2015
1x6mL	AdamTS13 Antigen Substrate	RQX32B0	31-Mar-2015
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL32B0	31-Mar-2015
1x6mL	AdamTS13 Antigen Stop Solution	RQK32B0	31-Mar-2015
1x80mL	Wash buffer Concentrate	RNA34B0	30-Sep-2015
1x90mL	Incubations buffer	RNB33B0	30-Sep-2015
1x0,5ml	Adamts 13 Calibrator 1	1UT32B0	30-Jun-2015
1x0,5ml	Adamts 13 Calibrator 2	2UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 3	3UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 4	4UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 5	6UT31B0	31-Mar-2015
1x0,5ml	Adamts 13 High Control	RUY32B0	31-Jul-2015
1x0,5ml	Adamts 13 Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RUY32B0 activity	46.3 – 77.2 %	63.13	OK
	High control RUY32B0 antigen	0.63 – 1.05 µg/mL	0.83	OK
	Low control RUZ32B0 activity	14.0 – 23.4 %	22.25	OK
	Low control RUZ32B0 antigen	0.19 – 0.32 µg/mL	0.25	OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	>250	448.1	OK
	Gain: 59 Antigen ΔRFU Cal1-Cal5	>250	344	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK

Durchführung von
Prepared by

Fikret NASUFI

F. Nasufi

20 - NOV - 2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER

[Signature]

21. NOV. 2013
Datum / Date: