

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRQ00AZ
---	---	----------

TECHNOZYM ADAMTS-13 ELISA Kit 2x48T.

REF 5450501

LOT RQ42B00.01

 28 - Feb - 2015

Inhalt Content	Reagenz Reagent	LOT	
1x48T	Adamts13 Plate	RQS41B0	29-Feb-2016
1x0.3mL	Adamts13 Conjugate	RQV31B0	28-Feb-2015
1x3mL	AdamTS13 Activity Substrate	RQW33B0	30-Nov-2015
1x6mL	AdamTS13 Antigen Substrate	RQX33B0	31-Aug-2015
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL33B0	31-Aug-2015
1x6mL	AdamTS13 Antigen Stop Solution	RQK33B0	31-Aug-2015
1x80mL	Wash buffer Concentrate	RNA35B0	31-Dec-2015
1x90mL	Incubations buffer	RNB41B0	31-Jan-2016
1x0,5ml	Adamts 13 Calibrator 1	1UT32B0	30-Jun-2015
1x0,5ml	Adamts 13 Calibrator 2	2UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 3	3UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 4	4UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 5	6UT32B0	30-Jun-2015
1x0,5ml	Adamts 13 High Control	RUY32B0	31-Jul-2015
1x0,5ml	Adamts 13 Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability	High control RUY32B0 activity	46.3– 77.2 %	58.8	OK
	High control RUY32B0 antigen	0.63 – 1.05 µg/mL	0.78	OK
	Low control RUZ32B0 activity	14.0 – 23.4 %	17.66	OK
	Low control RUZ32B0 antigen	0.19 – 0.32 µg/mL	0.25	OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	≥250	499.2	OK
	Gain: 61 Antigen ΔRFU Cal1-Cal5	≥250	256	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included	Ja / Yes	Ja / Yes	OK	

Durchführung von
Prepared by

Fikret NASUEI *Fikret Nasuei*

06. Mai. 2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA *Michael Riha*

06-MAY-2014
Datum / Date: