



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRQ00AZ

TECHNOZYM ADAMTS-13 ELISA Kit 2x48T.

REF 5450501

LOT RQ51B00.01

⌚ 30-Apr-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
2x48T	Adamts13 Plate	RQS42B0	31-Oct-2016
1x0.3mL	Adamts13 Conjugate	RQV32B1	30-Nov-2016
2x3mL	AdamTS13 Activity Substrate	RQW41B0	30-Apr-2016
1x6mL	AdamTS13 Antigen Substrate	RQX33B1	31-Dec-2016
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL33B1	31-Dec-2016
1x6mL	AdamTS13 Antigen Stop Solution	RQK33B1	31-Dec-2016
1x80mL	Wash buffer Concentrate	RNA43B0	31-Dec-2016
1x90mL	Incubations buffer	RNB51B0	28-Feb-2017
1x0.5ml	Adamts 13 Calibrator 1	1UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 2	2UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 3	3UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 4	4UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 5	6UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 High Control	RUY41B0	30-Jun-2016
1x0.5ml	Adamts 13 Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RUY41B0 activity	49.36 – 82.26 %	72.75	OK
	High control RUY41B0 antigen	0.64 – 1.06 µg/mL	0.97	OK
	Low control RUZ41B0 activity	11.00 – 25.68 %	19.76	OK
	Low control RUZ41B0 antigen	0.13 – 0.30 µg/mL	0.24	OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	≥250	476.2	OK
	Gain: 59 Antigen ΔRFU Cal1-Cal5	≥250	251	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

10-Apr-2015
Datum / Date:

Freigabe von
Approved by

Michael RIHA *Michael Riha*

10-APR-2015
Datum / Date: