


	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRO00AZ
---	---	----------

TECHNOZYM[®] ADAMTS-13 ELISA Kit 2x48T.

REF 5450501
LOT RQ61B00.01
 30-Nov-2016

Inhalt Content	Reagenz Reagent	LOT	
2x48T	Adamts13 Plate	RQS51B0	31-Jul-2017
1x0.3mL	Adamts13 Conjugate	RQV32B1	30-Nov-2016
1x3mL	AdamTS13 Activity Substrate	RQW41B1	31-Dec-2017
1x6mL	AdamTS13 Antigen Substrate	RQX33B1	31-Dec-2016
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL33B1	31-Dec-2016
1x6mL	AdamTS13 Antigen Stop Solution	RQK33B1	31-Dec-2016
1x80mL	Wash buffer Concentrate	RNA52B0	31-Oct-2017
1x90mL	Incubations buffer	RNB52B0	31-Jul-2017
1x0.5ml	Adamts 13 Calibrator 1	1UT53B0	31-Jul-2017
1x0.5ml	Adamts 13 Calibrator 2	2UT53B0	31-Jul-2017
1x0.5ml	Adamts 13 Calibrator 3	3UT53B0	31-Jul-2017
1x0.5ml	Adamts 13 Calibrator 4	4UT53B0	31-Jul-2017
1x0.5ml	Adamts 13 Calibrator 5	6UT53B0	31-Jul-2017
1x0.5ml	Adamts 13 High Control	RUY53B0	31-Jul-2017
1x0.5ml	Adamts 13 Low Control	RUZ53B0	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RUY53B0 activity	0.32 – 0.54 IU/mL	0.47	OK
	High control RUY53B0 antigen	0.32 – 0.54 IU/mL	0.41	OK
	Low control RUZ53B0 activity	0.07 – 0.17 IU/mL	0.13	OK
	Low control RUZ53B0 antigen	0.05 – 0.13 IU/mL	0.13	OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	≥200	431.1	OK
	Gain: 60 Antigen ΔRFU Cal1-Cal5	≥200	223	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRO00AZ
---	---	----------

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

25-Feb-2016
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

25-Feb-2016
Datum / Date: