



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRO00AZ

TECHNOZYM ADAMTS-13 ELISA Kit 1x48T.

REF 5450551

LOT RO32B00.01

 31-Mar-2015

Inhalt Content	Reagenz Reagent	LOT	
1x48T	Adamts13 Plate	RQS33B0	30-Jun-2015
1x0.3mL	Adamts13 Conjugate	RQV32B0	31-Aug-2015
1x3mL	AdamTS13 Activity Substrate	RQW31B0	31-Mar-2015
1x6mL	AdamTS13 Antigen Substrate	RQX32B0	31-Mar-2015
1x0.7mL	AdamTS13 Antigen Stable Peroxide	RQL32B0	31-Mar-2015
1x6mL	AdamTS13 Antigen Stop Solution	RQK32B0	31-Mar-2015
1x80mL	Wash buffer Concentrate	RNA33B0	31-May-2015
1x90mL	Incubations buffer	RNB32B0	30-Apr-2015
1x0,5ml	Adamts 13 Calibrator 1	1UT32B0	30-Jun-2015
1x0,5ml	Adamts 13 Calibrator 2	2UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 3	3UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 4	4UT32B0	31-Jul-2015
1x0,5ml	Adamts 13 Calibrator 5	6UT32B0	30-Jun-2015
1x0,5ml	Adamts 13 High Control	RUY32B0	31-Jul-2015
1x0,5ml	Adamts 13 Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RUY32B0 activity	46.3– 77.2 %	66.7 OK
	High control RUY32B0 antigen	0.63 – 1.05 µg/mL	0.83 OK
	Low control RUZ32B0 activity	14.0 – 23.4 %	20.4 OK
	Low control RUZ32B0 antigen	0.19 – 0.32 µg/mL	0.26 OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	>250	500.4 OK
	Gain: 60 Antigen ΔRFU Cal1-Cal5	>250	413 OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included	Ja / Yes	Ja / Yes	OK

Durchführung von
Prepared by

Fikret NASUFI

F. Nasufi

27- Sep-2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER

Nikolaus Binder

30. SEP. 2013
Datum / Date: