



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQQT00AZ

TECHNOZYM ADAMTS-13 Calibrator Set 5x0.5mL

REF 5450561

LOT 6QT51B0.01

⌚ 30-Jun-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1x0.5ml	Adamts 13 Calibrator 1	1UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 2	2UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 3	3UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 4	4UT41B0	30-Jun-2016
1x0.5ml	Adamts 13 Calibrator 5	6UT41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RUY41B0 activity	49.36 – 82.26 %	65.7	OK
	High control RUY41B0 antigen	0.64 – 1.06 µg/mL	0.84	OK
	Low control RUZ41B0 activity	11.00 – 25.68 %	17.6	OK
	Low control RUZ41B0 antigen	0.13 – 0.30 µg/mL	0.18	OK
Plate: RQS42B0 Conjugate: RQV32B1	Gain: 130 Activity ΔRFU/min Cal1-Cal5	>250	338.5	OK
	Gain: 60 Antigen ΔRFU Cal1-Cal5	>250	304	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

04-Feb-2015
Datum / Date:

Freigabe von
Approved by

Michael RIHA *Michael Riha*

05-FEB-2015
Datum / Date: