



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQQT00AZ

TECHNOZYM ADAMTS-13 Control Set 2x0.5mL

REF 5450563
LOT 7QT51B0.01
⌚ 30-Jun-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
1x0,5ml	Adamts 13 High Control	RUY41B0	30-Jun-2016
1x0,5ml	Adamts 13 Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit / Usability	High control RUY41B0 activity	49.36 – 82.26 %	77.9	OK
	High control RUY41B0 antigen	0.64 – 1.06 µg/mL	1.02	OK
Plate: RQS42B0 Conjugate: RQV32B1	Low control RUZ41B0 activity	11.00 – 25.68 %	20.3	OK
	Low control RUZ41B0 antigen	0.13 – 0.30 µg/mL	0.24	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included	Ja / Yes	Ja / Yes	OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

04-Feb-2015
Datum / Date:

Freigabe von
Approved by

Michael RIHA *Michael Riha*

05-FEB-2015
Datum / Date: