

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQQT00AZ
---	---	----------

TECHNOZYM ADAMTS-13 Control Set 2x0.5mL

REF 5450563
LOT 7QT51B0.02
 30-Jun-2016

Inhalt Content	Reagenz Reagent	LOT	
1x0,5ml	Adamts 13 High Control	RUY41B0	30-Jun-2016
1x0,5ml	Adamts 13 Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability	High control RUY41B0 activity	49.36 – 82.26 %	71.6 OK
	High control RUY41B0 antigen	0.64 – 1.06 µg/mL	0.90 OK
Plate: RQS42B0 Conjugate: RQV32B1	Low control RUZ41B0 activity	11.00 – 25.68 %	18.9 OK
	Low control RUZ41B0 antigen	0.13 – 0.30 µg/mL	0.23 OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included	Ja / Yes	Ja / Yes	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

23-Apr-2015
Datum / Date:

**Freigabe von
Approved by**

Michael RIHA *Michael Riha*

23-APR-2015
Datum / Date: