



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRR00AZ

**TECHNOZYM ADAMTS-13 Antigen ELISA Kit 96T.**

**REF** 5450601

**LOT** RR32B00.01

30 – Apr- 2015

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 12	Adamts13 Ag Plate	RRS32B0	31-Jun-2015
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B0	31-Aug-2015
1x 80mL	Wash buffer Concentrate	RNA33B0	31-Mai-2015
1 x 90mL	Incubation buffer	RNB32B0	30-Apr-2015
1 x 12mL	TMB Substrate	RKN32B0	31-Mai-2015
1 x 12mL	Stop solution	RKK32B0	30-Jun-2015
1 x 0,5mL	Adamts13 Ag Calibrator 1	1UT32B0	30-Jun-2015
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT32B0	31-Jul-2015
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT32B0	31-Jul-2015
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT32B0	31-Jul-2015
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT32B0	30-JUN-2015
1 x 0.5mL	Adamts13 Ag High Control	RUY32B0	31-Jul-2015
1 x 0.5mL	Adamts13 Ag Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.1	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RUY32B0)	0.6– 1.0 µg/mL	0.74	OK
	Low Control (RUZ32B0)	0.18 – 0.3 µg/mL	0.29	OK

Durchführung von  
Prepared by

**Fikret NASUFI**

F. Nasufi

26-Sep-2013  
Datum / Date:

Freigabe von  
Approved by

**Nikolaus BINDER**

Nikolaus Binder

26. SEP. 2013

Datum / Date: