



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRR00AZ

TECHNOZYM ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601

LOT RR42B00.01

⌚ 31-Aug-2015

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 Ag Plate	RRS41B0	30-Sep-2016
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B0	31-Aug-2015
1x 80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1 x 90mL	Incubation buffer	RNB42B0	30-Apr-2016
1 x 12mL	TMB Substrate	RKN34B0	30-Nov-2015
1 x 12mL	Stop solution	RKK42B0	30-Sep-2016
1 x 0,5mL	Adamts13 Ag Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag High Control	RUY41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	0.85	*
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RUY41B0)	0.62 – 1.02 µg/mL	1.00	OK
	Low Control (RUZx410)	0.13 – 0.31 µg/mL	0.24	OK

*Sonderfreigabe

Durchführung von
Prepared by

Michael RIHA

6-Nov-2014
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

06-Nov-2014
Datum / Date: