




**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**


DQRR00AZ

**TECHNOZYM ADAMTS-13 Antigen ELISA Kit 96T.**

**REF** 5450601

**LOT** RR51B00.01

 30-Jun-2016

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 12	Adamts13 Ag Plate	RRS51B0	31-Mar-2017
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B1	30-Nov-2016
1x 80mL	Wash buffer Concentrate	RNA43B0	31-Dec-2016
1 x 90mL	Incubation buffer	RNB51B0	28-Feb-2017
1 x 12mL	TMB Substrate	RKN41B0	31-Aug-2016
1 x 12mL	Stop solution	RKK43B0	31-Dec-2016
1 x 0,5mL	Adamts13 Ag Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag High Control	RUY41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	0.83	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RUY41B0)	0.62 – 1.02 µg/mL	0.98	OK
	Low Control (RUZ41B0)	0.13 – 0.31 µg/mL	0.22	OK

Durchführung von  
Prepared by

Fikret NASUFI 

21-Apr-2015  
Datum / Date:

Freigabe von  
Approved by

Michael RIHA 

21-APR-2015  
Datum / Date: