



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRR00AZ

TECHNOZYM ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601

LOT RR52B00.01

30-Jun-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 Ag Plate	RRS51B0	31-Mar-2017
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B1	30-Nov-2016
1x 80mL	Wash buffer Concentrate	RNA51B0	30-Apr-2017
1 x 90mL	Incubation buffer	RNB51B0	28-Feb-2017
1 x 12mL	TMB Substrate	RKN51B0	30-Apr-2017
1 x 12mL	Stop solution	RKK43B0	31-Dec-2016
1 x 0,5mL	Adamts13 Ag Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag High Control	RUY41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.05	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (RUY41B0)	0.62 – 1.02 µg/mL	0.87	OK
	Low Control (RUZ41B0)	0.13 – 0.31 µg/mL	0.21	OK

Durchführung von
Prepared by

Fikret NASUFI
Fikret Nasufi

14-Jul-2015
Datum / Date:

Freigabe von
Approved by

[Signature]

14-Jun-2015
Datum / Date: