

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQRR00AZ
---	---	----------

**TECHNOZYM® ADAMTS-13 Antigen ELISA Kit 96T.**

REF      5450601  
LOT      RR53B00.01  
      30-Jun-2016

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 12	Adamts13 Ag Plate	RRS51B0	31-Mar-2017
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B1	30-Nov-2016
1x 80mL	Wash buffer Concentrate	RNA51B0	30-Apr-2017
1 x 90mL	Incubation buffer	RNB52B0	31-Jul-2017
1 x 12mL	TMB Substrate	RKN51B0	30-Apr-2017
1 x 12mL	Stop solution	RKK51B0	30-Jun-2017
1 x 0,5mL	Adamts13 Ag Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Calibrator 5	6UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag High Control	RUY41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Ag Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	0.90	*
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RUY41B0)	0.62 – 1.02 µg/mL	0.98	OK
	Low Control (RUZ41B0)	0.13 – 0.31 µg/mL	0.21	OK

\*Sonderfreigabe

Durchführung von  
Prepared by

*Fikret NASUFI*

23. Oct - 2015  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA

23 OCT 2015  
Datum / Date: