



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRU00AZ

TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF 5450701

LOT RU3CB00.01

28 - Feb - 2015

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	335	28-Feb-2015
1 x 12mL	Adamts13 Activity HRP Conjugated antibody	335	28-Feb-2015
1x 53mL	Wash buffer concentrate	335	28-Feb-2015
1 x 30mL	Reaction buffer	335	28-Feb-2015
1 x 12mL	Color Reagent	335	28-Feb-2015
1 x 12mL	Stop solution	335	28-Feb-2015
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	335	28-Feb-2015
1 x	Sample Dilution MTP	335	28-Feb-2015
1 x 0.5ml	Adamts13 Activity Calibrator 1	1UT32B0	30-Jun-205
1 x 0.5ml	Adamts13 Activity Calibrator 2	2UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 3	3UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 4	4UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 5	5UT31B0	31-Mar-2015
1 x 0.5 mL	Adamts13 Activity Calibrator 6	6UT32B0	30-Jun-205
1 x 0.5ml	Adamts13 Activity High Control	RUY32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Low Control	RUZ32B0	31 Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.76	OK
	OD Cal 6	< 0.2	0.10	OK
	High Control (RUY32B0)	50.4% - 84.0%	76.25	OK
	Low Control (RUZ32B0)	17.2% - 28.6%	25.3	OK

Durchführung von
Prepared by

Fikret NASUFI

F. Nasufi

30-Sep-2013
Datum / Date:

Freigabe von
Approved by

Nikolaus BINDER

Nikolaus Binder

30. SEP. 2013
Datum / Date: