



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRU00AZ

TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF 5450701
LOT RU41B00.02
 30 - Jun - 2015

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	355	30-Jun-2015
1 x 12mL	Adamts13 Activity HRP Conj antibody	355	30-Jun-2015
1x 53mL	Wash buffer concentrate	355	30-Jun-2015
1 x 30mL	Reaction buffer	355	30-Jun-2015
1 x 12mL	Color Reagent	355	30-Jun-2015
1 x 12mL	Stop solution	355	30-Jun-2015
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	355	30-Jun-2015
1 x	Sample Dilution MTP	355	30-Jun-2015
1 x 0.5ml	Adamts13 Activity Calibrator 1	1UT32B0	30-Jun-2015
1 x 0.5ml	Adamts13 Activity Calibrator 2	2UT32B0	31-Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 3	3UT32B0	31-Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 4	4UT32B0	31-Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 5	5UT32B0	31-Jul-2015
1 x 0.5 ml	Adamts13 Activity Calibrator 6	6UT32B0	30-Jun-2015
1 x 0.5ml	Adamts13 Activity High Control	RUY32B0	31-Jul-2015
1 x 0.5ml	Adamts13 Activity Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	2.14	OK
	OD Cal 6	< 0.2	0.06	OK
	High Control (RUY32B0)	50.4% - 84.0%	77.7	OK
	Low Control (RUZ32B0)	17.2% - 28.6%	24.3	OK

Durchführung von
Prepared by

Fikret NASUFI

Fikret Nasufi

22-JAN-2014

Datum / Date:

Freigabe von
Approved by

Michael RIHA

Michael Riha

22-JAN-2014

Datum / Date: