



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRU00AZ

## TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF      5450701  
LOT      RU42B00.01  
      30 - Jun - 2015

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	405	31-Aug-2015
1 x 12mL	Adamts13 Activity HRP Conj antibody	405	31-Aug-2015
1x 53mL	Wash buffer concentrate	405	31-Aug-2015
1 x 30mL	Reaction buffer	405	31-Aug-2015
1 x 12mL	Color Reagent	405	31-Aug-2015
1 x 12mL	Stop solution	405	31-Aug-2015
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	405	31-Aug-2015
1 x	Sample Dilution MTP	405	31-Aug-2015
1 x 0.5ml	Adamts13 Activity Calibrator 1	1UT32B0	30-Jun-205
1 x 0.5ml	Adamts13 Activity Calibrator 2	2UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 3	3UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 4	4UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 5	5UT32B0	31-Jul-2015
1 x 0.5 ml	Adamts13 Activity Calibrator 6	6UT32B0	30-Jun-205
1 x 0.5ml	Adamts13 Activity High Control	RUY32B0	31-Jul-2015
1 x 0.5ml	Adamts13 Activity Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.37    OK
	OD Cal 6	< 0.2	0.05    OK
	High Control (RUY32B0)	50.4% - 84.0%	78.97    OK
	Low Control (RUZ32B0)	17.2% - 28.6%	28.47    OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

05-Mär-2014  
Datum / Date:

Freigabe von  
Approved by

Michael RIHA *Michael Riha*

05-Mär-2014  
Datum / Date: