




| | | |
|---|---|----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQRU00AZ |
|---|---|----------|

TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF 5450701
LOT RU46B00.01
 30 - Jun - 2015

| Inhalt/ Content | Reagenz/Reagent | LOT |  |
|--------------------|-------------------------------------|---------|---|
| 1 x 12 | Adamts13 Activity Strips | 411 | 30-Sep-2015 |
| 1 x 12mL | Adamts13 Activity HRP Conj antibody | 411 | 30-Sep-2015 |
| 1x 53mL | Wash buffer concentrate | 411 | 30-Sep-2015 |
| 1 x 30mL | Reaction buffer | 411 | 30-Sep-2015 |
| 1 x 12mL | Color Reagent | 411 | 30-Sep-2015 |
| 1 x 12mL | Stop solution | 411 | 30-Sep-2015 |
| 2 x 6mL | ADAMTS-13 Activity vWF73 Substrate | 411 | 30-Sep-2015 |
| 1 x | Sample Dilution MTP | 411 | 30-Sep-2015 |
| 1 x 0.5ml | Adamts13 Activity Calibrator 1 | 1UT32B0 | 30-Jun-205 |
| 1 x 0.5ml | Adamts13 Activity Calibrator 2 | 2UT32B0 | 31 Jul-2015 |
| 1 x 0.5ml | Adamts13 Activity Calibrator 3 | 3UT32B0 | 31 Jul-2015 |
| 1 x 0.5ml | Adamts13 Activity Calibrator 4 | 4UT32B0 | 31 Jul-2015 |
| 1 x 0.5ml | Adamts13 Activity Calibrator 5 | 5UT32B0 | 31-Jul-2015 |
| 1 x 0.5 ml | Adamts13 Activity Calibrator 6 | 6UT32B0 | 30-Jun-205 |
| 1 x 0.5ml | Adamts13 Activity High Control | RUY32B0 | 31-Jul-2015 |
| 1 x 0.5ml | Adamts13 Activity Low Control | RUZ32B0 | 31-Jul-2015 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| Verwendbarkeit Usability | Top OD Cal 1 | > 1 | 1.70 | OK |
| | OD Cal 6 | < 0.2 | 0.05 | OK |
| | High Control (RUY32B0) | 50.4% - 84.0% | 70.08 | OK |
| | Low Control (RUZ32B0) | 17.2% - 28.6% | 24.04 | OK |

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

03-Apr-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA *Michael Riha*

03-APR-2014
Datum / Date: