



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRU00AZ

**TECHNOZYM ADAMTS-13 Activity ELISA 96T.**

REF 5450701  
 LOT RU4BB00.01  
 ⌚ 30 - Jun - 2015

Inhalt/ Content	Reagenz/Reagent	LOT	⌚
1 x 12	Adamts13 Activity Strips	425	31-Dec-2015
1 x 12mL	Adamts13 Activity HRP Conj antibody	425	31-Dec-2015
1x 53mL	Wash buffer concentrate	425	31-Dec-2015
1 x 30mL	Reaction buffer	425	31-Dec-2015
1 x 12mL	Color Reagent	425	31-Dec-2015
1 x 12mL	Stop solution	425	31-Dec-2015
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	425	31-Dec-2015
1 x	Sample Dilution MTP	425	31-Dec-2015
1 x 0.5ml	Adamts13 Activity Calibrator 1	1UT32B0	30-Jun-205
1 x 0.5ml	Adamts13 Activity Calibrator 2	2UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 3	3UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 4	4UT32B0	31 Jul-2015
1 x 0.5ml	Adamts13 Activity Calibrator 5	5UT32B0	31-Jul-2015
1 x 0.5 ml	Adamts13 Activity Calibrator 6	6UT32B0	30-Jun-205
1 x 0.5ml	Adamts13 Activity High Control	RUY32B0	31-Jul-2015
1 x 0.5ml	Adamts13 Activity Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.15	OK
	OD Cal 6	< 0.2	0.05	OK
	High Control (RUY32B0)	50.4% - 84.0%	81.87	OK
	Low Control (RUZ32B0)	17.2% - 28.6%	27.08	OK

Durchführung von  
Prepared by

Fikret NASUFI

*Fikret Nasufi*

14-Jul-2014  
Datum / Date:

Freigabe von  
Approved by

Michael RIHA

*Michael Riha*

14-Jul-2014  
Datum / Date: