




**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**


DQRU00AZ

TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF 5450701

LOT RU4GB00.01

 29-Feb-2016

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	436	29-Feb-2016
1 x 12mL	Adamts13 Activity HRP Conj antibody	436	29-Feb-2016
1x 53mL	Wash buffer concentrate	436	29-Feb-2016
1 x 30mL	Reaction buffer	436	29-Feb-2016
1 x 12mL	Color Reagent	436	29-Feb-2016
1 x 12mL	Stop solution	436	29-Feb-2016
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	436	29-Feb-2016
1 x	Sample Dilution MTP	436	29-Feb-2016
1 x 0.5ml	Adamts13 Activity Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 5	5UT41B0	30-Jun-2016
1 x 0.5 ml	Adamts13 Activity Calibrator 6	6UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity High Control	RUY41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.25	OK
	OD Cal 6	< 0.2	0.03	OK
	High Control (RUY41B0)	63.20 – 105.31%	88.77	OK
	Low Control (RUZ41B0)	13.80 – 32.21%	24.34	OK

Durchführung von
Prepared by

Fikret NASUFI
Fikret Nasufi

27-Oct-2014
Datum / Date:

Freigabe von
Approved by

Michael RIHA
Michael Riha

27-OCT-2014
Datum / Date: