

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQRU00AZ
---	---	----------

**TECHNOZYM ADAMTS-13 Activity ELISA 96T.**

REF      5450701  
LOT      RU52B00.01  
      30-Jun-2016

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	455	30-Jun-2016
1 x 12mL	Adamts13 Activity HRP Conj antibody	455	30-Jun-2016
1x 53mL	Wash buffer concentrate	455	30-Jun-2016
1 x 30mL	Reaction buffer	455	30-Jun-2016
1 x 12mL	Color Reagent	455	30-Jun-2016
1 x 12mL	Stop solution	455	30-Jun-2016
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	455	30-Jun-2016
1 x	Sample Dilution MTP	455	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Calibrator 5	5UT41B0	30-Jun-2016
1 x 0.5 ml	Adamts13 Activity Calibrator 6	6UT41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity High Control	RUY41B0	30-Jun-2016
1 x 0.5ml	Adamts13 Activity Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.01	OK
	OD Cal 6	< 0.2	0.03	OK
	High Control (RUY41B0)	63.20 – 105.31%	102.66	OK
	Low Control (RUZ41B0)	13.80 – 32.21%	18.51	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

10-Feb-2015  
Datum / Date:

Freigabe von  
Approved by

Michael RIHA *Michael Riha*

10-FEB-2015  
Datum / Date: