

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRU00AZ
---	---	----------

### TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF      5450701  
LOT      RU57B00.01  
      30-Jun-2016

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	515	31-Oct-2016
1 x 12mL	Adamts13 Activity HRP Conj antibody	515	31-Oct-2016
1x 53mL	Wash buffer concentrate	515	31-Oct-2016
1 x 30mL	Reaction buffer	515	31-Oct-2016
1 x 12mL	Color Reagent	515	31-Oct-2016
1 x 12mL	Stop solution	515	31-Oct-2016
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	515	31-Oct-2016
1 x	Sample Dilution MTP	515	31-Oct-2016
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity High Control	RUY41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.72	OK
	OD Cal 6	< 0.2	0.05	OK
	High Control (RUY41B0)	63.20 - 105.31%	101.58	OK
	Low Control (RUZ41B0)	13.80 - 32.21%	27.97	OK



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRU00AZ

Durchführung von  
Prepared by

Fikret NASUFI  
*Fikret Nasufi*

12-Jan-2015  
Datum / Date:

Freigabe von  
Approved by

*[Signature]*

12-Jan-2015  
Datum / Date: