

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRU00AZ
---	---	----------

TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF 5450701
LOT RU59B00.02
 30-Jun-2016

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	525	31-Dec-2016
1 x 12mL	Adamts13 Activity HRP Conj antibody	525	31-Dec-2016
1x 53mL	Wash buffer concentrate	525	31-Dec-2016
1 x 30mL	Reaction buffer	525	31-Dec-2016
1 x 12mL	Color Reagent	525	31-Dec-2016
1 x 12mL	Stop solution	525	31-Dec-2016
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	525	31-Dec-2016
1 x	Sample Dilution MTP	525	31-Dec-2016
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity High Control	RUY41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.54	OK
	OD Cal 6	< 0.2	0.03	OK
	High Control (RUY41B0)	63.20 – 105.31%	95.69	OK
	Low Control (RUZ41B0)	13.80 – 32.21%	25.68	OK


technoclone



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRU00AZ

Durchführung von
Prepared by



10-Jul-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI


10-Jul-2015
Datum / Date: