



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRU00AZ

TECHNOZYM ADAMTS-13 Activity ELISA 96T.


REF 5450701

LOT RU5BB00.02


 30-Jun-2016

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	530	31-Jan-2017
1 x 12mL	Adamts13 Activity HRP Conj antibody	530	31-Jan-2017
1x 53mL	Wash buffer concentrate	530	31-Jan-2017
1 x 30mL	Reaction buffer	530	31-Jan-2017
1 x 12mL	Color Reagent	530	31-Jan-2017
1 x 12mL	Stop solution	530	31-Jan-2017
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	530	31-Jan-2017
1 x	Sample Dilution MTP	530	31-Jan-2017
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity High Control	RUY41B0	30-Jun-2016
1 x 0.5mL	Adamts13 Activity Low Control	RUZ41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.41	OK
	OD Cal 6	< 0.2	0.02	OK
	High Control (RUY41B0)	63.20% - 105.31%	98.19	OK
	Low Control (RUZ41B0)	13.80% - 32.21%	23.30	OK

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQRU00AZ</p>
---	---	---

Durchführung von
Prepared by



24-Aug-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI


24-Aug-2015
Datum / Date: