



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRU00AZ

## TECHNOZYM ADAMTS-13 Activity ELISA 96T.

**REF** 5450701

**LOT** RU5EB00.01

31-Jan-2017

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	531	31-Jan-2017
1 x 12mL	Adamts13 Activity HRP Conj antibody	531	31-Jan-2017
1x 53mL	Wash buffer concentrate	531	31-Jan-2017
1 x 30mL	Reaction buffer	531	31-Jan-2017
1 x 12mL	Color Reagent	531	31-Jan-2017
1 x 12mL	Stop solution	531	31-Jan-2017
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	531	31-Jan-2017
1 x	Sample Dilution MTP	531	31-Jan-2017
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity High Control	RUY51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Low Control	RUZ51B0	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.55	OK
	OD Cal 6	< 0.2	0.02	OK
	High Control (RUY51B0)	0.55 – 0.91 IU/mL	0.69	OK
	Low Control (RUZ51B0)	0.13 – 0.29 IU/mL	0.22	OK



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRU00AZ

Durchführung von  
Prepared by

Michael KAFKA

A handwritten signature in black ink, appearing to read "Kafka", written over a horizontal line.

29-Sep-2015

Datum / Date:

Freigabe von  
Approved by

Alessandro BENIGNI

A handwritten signature in black ink, appearing to read "Alessandro Benigni", written over a horizontal line.

29-Sep-2015

Datum / Date: