



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRU00AZ

TECHNOZYM ADAMTS-13 Activity ELISA 96T.

REF 5450701

LOT RU5GB00.01

⌚ 30-Apr-2017

Inhalt/ Content	Reagenz/Reagent	LOT	⌚
1 x 12	Adamts13 Activity Strips	545	30-Apr-2017
1 x 12mL	Adamts13 Activity HRP Conj antibody	545	30-Apr-2017
1x 53mL	Wash buffer concentrate	545	30-Apr-2017
1 x 30mL	Reaction buffer	545	30-Apr-2017
1 x 12mL	Color Reagent	545	30-Apr-2017
1 x 12mL	Stop solution	545	30-Apr-2017
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	545	30-Apr-2017
1 x	Sample Dilution MTP	545	30-Apr-2017
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT5B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity High Control	RUY51B0	31-Jul-2017
1 x 0.5mL	Adamts13 Activity Low Control	RUZ51B0	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.25	OK
	OD Cal 6	< 0.2	0.02	OK
	High Control (RUY51B0)	0.55 – 0.91 IU/mL	0.59	OK
	Low Control (RUZ51B0)	0.13 – 0.29 IU/mL	0.20	OK



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRU00AZ

Durchführung von
Prepared by

Michael KAFKA

Handwritten signature of Michael Kafka in black ink, written over a horizontal line.

16-Nov-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUELI

Handwritten signature of Fikret Nasuefi in black ink, written over a horizontal line.

16-Nov-2015
Datum / Date: