




**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**


DQRU00AZ

TECHNOZYM® ADAMTS-13 Activity ELISA 96T.


REF 5450701

LOT RU5HB00.01

 31-May-2017

| Inhalt/ Content | Reagenz/Reagent | LOT |  |
|--------------------|-------------------------------------|---------|---|
| 1 x 12 | Adamts13 Activity Strips | 550 | 31-May-2017 |
| 1 x 12mL | Adamts13 Activity HRP Conj antibody | 550 | 31-May-2017 |
| 1x 53mL | Wash buffer concentrate | 550 | 31-May-2017 |
| 1 x 30mL | Reaction buffer | 550 | 31-May-2017 |
| 1 x 12mL | Color Reagent | 550 | 31-May-2017 |
| 1 x 12mL | Stop solution | 550 | 31-May-2017 |
| 2 x 6mL | ADAMTS-13 Activity vWF73 Substrate | 550 | 31-May-2017 |
| 1 x | Sample Dilution MTP | 550 | 31-May-2017 |
| 1 x 0.5mL | Adamts13 Activity Calibrator 1 | 1UT51B0 | 31-Jul-2017 |
| 1 x 0.5mL | Adamts13 Activity Calibrator 2 | 2UT51B0 | 31-Jul-2017 |
| 1 x 0.5mL | Adamts13 Activity Calibrator 3 | 3UT51B0 | 31-Jul-2017 |
| 1 x 0.5mL | Adamts13 Activity Calibrator 4 | 4UT51B0 | 31-Jul-2017 |
| 1 x 0.5mL | Adamts13 Activity Calibrator 5 | 5UT51B0 | 31-Jul-2017 |
| 1 x 0.5mL | Adamts13 Activity Calibrator 6 | 6UT51B0 | 31-Jul-2017 |
| 1 x 0.5mL | Adamts13 Activity High Control | RUY51B0 | 31-Jul-2017 |
| 1 x 0.5mL | Adamts13 Activity Low Control | RUZ51B0 | 31-Jul-2017 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| Verwendbarkeit Usability | Top OD Cal 1 | > 1 | 1.24 | OK |
| | OD Cal 6 | < 0.2 | 0.03 | OK |
| | High Control (RUY51B0) | 0.55 – 0.91 IU/mL | 0.63 | OK |
| | Low Control (RUZ51B0) | 0.13 – 0.29 IU/mL | 0.21 | OK |

| | | |
|---|---|---|
|  | <p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p> | <p style="text-align: center;">DQRU00AZ</p> |
|---|---|---|

Durchführung von
Prepared by

Michael KAFKA 

03-DEC-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUELI 

03-Dec-2015
Datum / Date: