



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ8UT0AZ.02

ADAMTS-13 Activity Calibrator Set 6x0.5mL

REF 5450761

LOT 8UT32B0.02

30 - Jun - 2015

Inhalt Content	Reagenz Reagent	LOT	
1X0.5mL	Adamts-13 Activity Calibrator 1	1UT32B0	30-Jun-2015
1X0.5mL	Adamts-13 Activity Calibrator 2	2UT32B0	31-Jul-2015
1X0.5mL	Adamts-13 Activity Calibrator 3	3UT32B0	31-Jul-2015
1X0.5mL	Adamts-13 Activity Calibrator 4	4UT32B0	31-Jul-2015
1X0.5mL	Adamts-13 Activity Calibrator 5	5UT32B0	31-Jul-2015
1x0.5mL	Adamts-13 Activity Calibrator 6	6UT32B0	30-Jun-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.79	OK
	Calibrator 6	< 0.2	0.104	OK
Verwendbarkeit Usability (tested with Adamts-13 Activity Strips 335)	High Control (RUY32B0)	50.0 – 84.0 %	74.84	OK
	Low Control (RUZ32B0)	17.2 – 28.6 %	26.94	OK

Durchführung von  
Prepared by

Fikret NASUFI

12-Nov-2013  
Datum / Date:

Freigabe von  
Approved by

Nikolaus BINDEF

12-Nov-2013  
Datum / Date: