




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ8UT0AZ.02

TECHNOZYM[®] ADAMTS-13 CAL Set 6x0.5mL

REF 5450761

LOT 8UT51B0.01

 30-Jun-2016

Inhalt Content	Reagenz Reagent	LOT	
1X0.5mL	Adamts-13 Activity Calibrator 1	1UT41B0	30-Jun-2016
1X0.5mL	Adamts-13 Activity Calibrator 2	2UT41B0	30-Jun-2016
1X0.5mL	Adamts-13 Activity Calibrator 3	3UT41B0	30-Jun-2016
1X0.5mL	Adamts-13 Activity Calibrator 4	4UT41B0	30-Jun-2016
1X0.5mL	Adamts-13 Activity Calibrator 5	5UT41B0	30-Jun-2016
1x0.5mL	Adamts-13 Activity Calibrator 6	6UT41B0	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.24	OK
	Calibrator 6	< 0.2	0.03	OK
Verwendbarkeit / Usability	High Control (RUY41B0)	63.20 – 105.31 %	83.65	OK
	Low Control (RUZ41B0)	13.80 – 32.21 %	21.27	OK

Durchführung von
Prepared by

Fikret NASUFI 

26-Mar-2015
Datum / Date:

Freigabe von
Approved by

Michael RIHA 

26-MAR-2015
Datum / Date: