

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8UT0AZ.02
---	---	-------------

TECHNOZYM® ADAMTS-13 Activity Calibrator Set 6x0.5mL

REF 5450761
LOT 8UT52B0.01
 31-Jul-2017

Inhalt Content	Reagenz Reagent	LOT	
1X0.5mL	Adamts-13 Activity Calibrator 1	1UT51B0	31-Jul-2017
1X0.5mL	Adamts-13 Activity Calibrator 2	2UT51B0	31-Jul-2017
1X0.5mL	Adamts-13 Activity Calibrator 3	3UT51B0	31-Jul-2017
1X0.5mL	Adamts-13 Activity Calibrator 4	4UT51B0	31-Jul-2017
1X0.5mL	Adamts-13 Activity Calibrator 5	5UT51B0	31-Jul-2017
1x0.5mL	Adamts-13 Activity Calibrator 6	6UT51B0	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.36	OK
	Calibrator 6	< 0.2	0.03	OK
Verwendbarkeit / Usability	High Control (RUY51B0)	0.55 – 0.91 IU/mL	0.79	OK
	Low Control (RUZ51B0)	0.13 – 0.29 IU/mL	0.26	OK

Durchführung von
Prepared by

Fikret NASUFI

05-Nov-2015
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

05-Nov-2015
Datum / Date: