technoclone TC	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7UT0AZ.02

ADAMTS-13 Activity Control Set 2x0.5mL

REF 5450763

LOT 7UT31B0.02

⌛ 31 - Mar - 2015

Inhalt Content	Reagenz Reagent	LOT	⌛
1X0.5mL	Adamts-13 Activity High Control	RUY31B0	31-Mar-2015
1X0.5mL	Adamts-13 Low Control	RUZ31B0	31-Mar-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability (tested with Adamts-13 Activity Strips 321)	High Control (RUY31B0)	53.8 – 89.7 %	78.9	OK
	Low Control (RUZ31B0)	12.4 – 28.9 %	24.9	OK

Durchführung von
Prepared by

F. Nasufi

16-Jul-2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK

16. JULI 2013
Datum / Date: