

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ7UT0AZ.02
---	---	-------------

**ADAMTS-13 Activity Control Set 2x0.5mL**

**REF**            5450763  
**LOT**            7UT32B0.01  
        31 - Jul - 2015

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1X0.5mL	Adamts-13 Activity High Control	RUY32B0	31-Jul-2015
1X0.5mL	Adamts-13 Low Control	RUZ32B0	31-Jul-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability (tested with Adamts-13 Activity Strips 335)	High Control (RUY32B0)	50.4 – 84.0 %	74.84	OK
	Low Control (RUZ32B0)	17.2 – 28.9 %	26.94	OK

Durchführung von  
Prepared by

**Fikret NASUFI**

*F. Nasufi*

12-Nov-2013  
Datum / Date:

Freigabe von  
Approved by

**Nikolaus BINDER**

*[Signature]*

12-Nov-2013  
Datum / Date: