

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9900AZ.02 LOT 9953B00
---	---	---------------------------------------

Ceveron[®] System Solution, 750mL

REF	9820210
LOT	9953B00
	02-Mar-2015
	31-Mar-2020

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	2.6 – 3.1	2.7	OK
Volumen volume	> 750.0 mL	780.0	OK
Sterilität sterility	0 CFU/mL	0	OK

Durchführung von
Prepared by


Monika WEINBAUER

05-Mar-2015
Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

05-Mar-2015
Datum / Date:

Ceveron[®] alpha


S. SCHWINGENSCHLÖGL

05-Mar-2015
Datum / Date: