



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ9900AZ.01

LOT 995GB00

Ceveron[®] System Solution, 750mL

REF 9820210
LOT 995GB00
17-Nov-2015
30-Nov-2020

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	2.6 – 3.1	2.8	OK
Volumen volume	> 750.0 mL	790.0	OK
Sterilität sterility	0 CFU/mL	0	OK

Durchführung von
Prepared by

Lavdrim NASUFI

23-Nov-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

23-NOV-2015
Datum / Date:

Ceveron[®] alpha

S. SCHWINGENSCHLÖGL

23-NOV-2015
Datum / Date: